

158431  
5115F1401

REGION V  
UNITED STATES ENVIRONMENTAL PROTECTION AGENCY

DATE:

SUBJECT: Review of Region V Contractor Data; Received for Review on

FROM: Curtis Ross, Director  
Central Regional Laboratory

TO: Data User:

TAT (Lisa Parachuri)

We have reviewed the data for the following Case(s):

Site Name: SAUGET VIAR Case No.: 801

EPA Data Set No.: SF1401 Decision Unit: Y306

CRL Laboratory Numbers: 82WT06502, 82WT06504, 82WT06506

VIAR Traffic Numbers: E 1202, 1203, 1204

Person-hours required for review: 2.0

Following are our findings:

- (1) Volatile matrix spikes average recovery is 154% - higher than ~~usual~~ the usually acceptable value.
- (2) All other Quality control appears to be in compliance.

☒ Data are acceptable for use.

☐ Data are unacceptable for use.

☐ Data are preliminary - this case has been forwarded to Dr. Alfred Haerberer, EPA Support Services, for review - pending reply.

cc: Dr. Alfred Haerberer